



National Centre for Physics
Detailed specification Document

Annex - A - 176

S#	Title of the Project	Equipment required for leak testing jobs		
		Specification	A/U	QTY
1.	Rotary vane pump	<ul style="list-style-type: none">• KRL Made or China made• Pumping speed: 36-40 m³/h• Inlet port size: ISO 40 KF• Double stage	Nos	01

Note: All the components installed should be compatible with the system and each other.
Please tick the relevant box

Installation Testing Commissioning Training
 Warranty requirement Any other please specify

✓

✓